
Midmark M11 - v1.0.0
Total Cycles: 3607

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 178.4 F 0.8
4:00 187.2 F 4.8
6:00 208.3 F 10.0
8:00 222.2 F 6.5
10:00 231.5 F 9.6
12:00 236.3 F 9.3
14:00 242.8 F 12.2
16:00 240.4 F 10.4
18:00 253.2 F 16.8
20:00 264.0 F 23.3
22:00 271.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.8 F 28.9
2:00 271.7 F 28.2
3:00 272.4 F 28.9
4:00 272.4 F 28.7

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 225.6 F 0.0
10:00 234.7 F 0.0
15:00 228.6 F 0.0
20:00 222.8 F 0.0
25:00 224.8 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

08/02/2023 12:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____