
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008034
11:04 07/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:37
STERILIZING 3:36
136.0 C 220kPa 3:36
136.4 C 222kPa 4:06
136.1 C 221kPa 4:36
136.2 C 221kPa 5:06
136.3 C 223kPa 5:36
136.4 C 222kPa 6:06
136.4 C 224kPa 6:36
136.4 C 224kPa 7:06
136.4 C 224kPa 7:06
VENTING 7:06
AIR DRYING 7:59
CYCLE COMPLETE 37:59

Digital Signature #
49A6DBA9FC26C06A
