

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5184

Name: RITTER 1  
Sterilizer ID: V2178198  
12/12/2024 03:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.9 F 0.0  
2:00 95.3 F 0.0  
4:00 118.6 F 0.0  
6:00 139.2 F 0.1  
8:00 166.5 F 0.2  
10:00 187.7 F 0.5  
12:00 204.2 F 3.5  
14:00 222.1 F 6.6  
16:00 228.0 F 10.2  
18:00 234.6 F 9.2  
20:00 238.7 F 10.7  
22:00 235.2 F 8.0  
24:00 243.4 F 11.6  
26:00 246.0 F 13.0  
28:00 257.5 F 19.1  
30:00 267.0 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.2  
1:00 272.1 F 28.6  
2:00 272.7 F 29.0  
3:00 272.5 F 28.7  
4:00 272.1 F 28.5

Min 271.5 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.9 F 0.0  
5:00 224.8 F 0.0  
10:00 215.1 F 0.0  
15:00 215.3 F 0.0  
20:00 236.3 F 0.0  
25:00 231.4 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:44  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:10:14  
CYCLE COMPLETE

12/12/2024 05:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_