
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005507
11:36 10/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:06
STERILIZING 3:04
136.0 C 220kPa 3:04
136.3 C 223kPa 3:34
136.4 C 222kPa 4:04
136.4 C 219kPa 4:34
136.4 C 224kPa 5:04
136.2 C 220kPa 5:34
136.4 C 223kPa 6:04
136.4 C 223kPa 6:34
136.4 C 223kPa 6:34
VENTING 6:34
AIR DRYING 7:22
CYCLE COMPLETE 31:50

Digital Signature #
460F9AE5FA3B34A2
