
Midmark M11 - v1.0.0
Total Cycles: 4988

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2024 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.5 F 0.0
2:00 164.4 F 0.9
4:00 186.8 F 4.7
6:00 204.8 F 9.5
8:00 216.4 F 7.5
10:00 229.1 F 8.4
12:00 241.1 F 13.4
14:00 240.2 F 10.8
16:00 237.6 F 9.2
18:00 251.8 F 16.0
20:00 263.6 F 22.9
22:00 271.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.7
2:00 271.6 F 28.2
3:00 272.6 F 28.8
4:00 271.8 F 28.1

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 214.4 F 0.0
10:00 207.8 F 0.0
15:00 211.8 F 0.0
20:00 216.4 F 0.0
25:00 220.7 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

10/10/2024 04:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____