

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2806

Name: RITTER2  
Sterilizer ID: V2313228  
08/28/2023 03:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 181.5 F 0.9  
4:00 204.4 F 5.5  
6:00 220.4 F 10.2  
8:00 227.7 F 7.3  
10:00 237.5 F 11.8  
12:00 237.1 F 9.9  
14:00 237.5 F 9.0  
16:00 250.2 F 15.7  
18:00 261.9 F 22.4  
20:00 272.4 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.7 F 29.5  
2:00 271.6 F 28.7  
3:00 272.4 F 29.6  
4:00 272.4 F 29.1

Min 271.2 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 210.6 F 0.0  
10:00 216.9 F 0.0  
15:00 223.3 F 0.0  
20:00 223.2 F 0.0  
25:00 228.9 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:53  
HEATING: 20:28  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:59:42  
CYCLE COMPLETE

08/28/2023 04:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_