
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000909
16:30 04/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:51
STERILIZING 4:00
136.0 C 228kPa 4:00
136.4 C 232kPa 4:31
136.3 C 232kPa 5:01
136.4 C 233kPa 5:31
136.4 C 233kPa 6:01
136.2 C 231kPa 6:31
136.1 C 230kPa 7:01
136.4 C 233kPa 7:31
136.4 C 233kPa 7:31
VENTING 7:31
AIR DRYING 8:26
CYCLE COMPLETE 12:43

Digital Signature #
CADACB45F874C0AC
