
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001549
12:38 11/01/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:42
PRESSURIZING 6:20
STERILIZING 7:50
136.0 C 220kPa 7:50
136.1 C 220kPa 8:20
136.4 C 220kPa 8:50
136.4 C 221kPa 9:20
136.4 C 223kPa 9:50
136.3 C 221kPa 10:20
136.1 C 220kPa 10:50
136.4 C 221kPa 11:20
VENTING 11:20
AIR DRYING 12:21
CYCLE COMPLETE 42:21

Digital Signature #
01990D1C269CCA14
