

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1285

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2022 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.9 F 0.0  
2:00 142.2 F 1.4  
4:00 151.6 F 4.1  
6:00 178.9 F 8.5  
8:00 207.4 F 7.7  
10:00 229.8 F 8.7  
12:00 239.0 F 9.7  
14:00 241.9 F 12.5  
16:00 241.6 F 11.4  
18:00 255.0 F 18.4  
20:00 266.8 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.0 F 28.9  
2:00 272.6 F 29.3  
3:00 272.7 F 29.7  
4:00 271.7 F 28.9

Min 271.5 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 229.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:02  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 2:46  
DRYING: 4:26  
TOTAL CYCLE: 00:33:39

07/12/2022 02:46 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_