
Midmark M11 - v1.0.0
Total Cycles: 2970

Name: RITTER 1
Sterilizer ID: V2178198
03/03/2023 11:54 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	109.2 F	0.0
2:00	111.0 F	0.0
4:00	135.3 F	0.0
6:00	169.5 F	0.0
8:00	198.6 F	0.0
10:00	212.0 F	0.1
12:00	227.7 F	5.4
14:00	239.1 F	10.0
16:00	238.1 F	8.8
18:00	246.5 F	12.4
20:00	246.4 F	13.0
22:00	245.5 F	12.6
24:00	260.6 F	20.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 252F
CHAMBER PRESS 16.0PSI
** ITEMS NOT STERILE **

03/03/2023 12:19 PM