

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000139  
8:53 11/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:21  
PRESSURIZING 3:23  
STERILIZING 4:31  
136.0 C 221kPa 4:31  
135.9 C 221kPa 5:01  
136.4 C 224kPa 5:31  
136.4 C 221kPa 6:01  
136.4 C 225kPa 6:31  
136.3 C 223kPa 7:01  
136.2 C 223kPa 7:31  
136.4 C 224kPa 8:01  
136.4 C 224kPa 8:01  
VENTING 8:01  
AIR DRYING 8:56  
CYCLE COMPLETE 23:56  
-----

Digital Signature #  
4DD7159591788770  
-----