
Midmark M11 - v1.0.5
Total Cycles: 2328

Name: RITTER2
Sterilizer ID: V2313228
05/03/2023 07:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.3 F 0.0
2:00 78.4 F 0.0
4:00 82.0 F 0.0
6:00 99.6 F 1.2
8:00 132.1 F 3.7
10:00 162.3 F 6.6
12:00 187.1 F 10.1
14:00 217.4 F 8.1
16:00 219.1 F 7.3
18:00 227.1 F 11.4
20:00 229.8 F 7.5
22:00 241.3 F 12.8
24:00 239.8 F 11.0
26:00 251.5 F 16.9
28:00 261.9 F 22.9
30:00 271.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.8
1:00 272.4 F 30.6
2:00 272.1 F 29.8
3:00 272.4 F 30.6
4:00 272.3 F 29.9

Min 271.7 F 29.7
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 212.6 F 0.0
10:00 214.2 F 0.0
15:00 200.4 F 0.0
20:00 198.1 F 0.0
25:00 195.6 F 0.0
30:00 193.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:17
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

05/03/2023 08:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____