
Midmark M11 - v1.0.0
Total Cycles: 4683

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2024 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 74.3 F 0.3
4:00 98.2 F 1.3
6:00 128.1 F 3.5
8:00 155.6 F 6.4
10:00 179.9 F 9.9
12:00 199.9 F 13.4
14:00 213.8 F 4.3
16:00 223.0 F 7.5
18:00 230.4 F 11.2
20:00 227.8 F 6.1
22:00 240.1 F 11.3
24:00 235.3 F 8.5
26:00 248.8 F 14.7
28:00 259.8 F 20.8
30:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.7 F 29.1
2:00 271.4 F 28.2
3:00 272.3 F 29.2
4:00 272.4 F 28.7

Min 271.1 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 232.5 F 0.0
10:00 221.7 F 0.0
15:00 229.3 F 0.0
20:00 214.4 F 0.0
25:00 220.4 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:09:08
CYCLE COMPLETE

06/27/2024 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____