
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002023
11:37 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:59
STERILIZING 3:01
136.4 C 225kPa 3:01
136.3 C 225kPa 3:31
136.3 C 224kPa 4:01
136.4 C 223kPa 4:31
136.4 C 226kPa 5:01
136.4 C 225kPa 5:31
136.0 C 221kPa 6:01
136.4 C 224kPa 6:31
136.4 C 224kPa 6:31
VENTING 6:31
AIR DRYING 7:23
CYCLE COMPLETE 14:37

Digital Signature #
D6EAF63119E80D9A
