
Midmark M11 - v1.0.0
Total Cycles: 5191

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2024 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 69.5 F 0.0
4:00 81.1 F 0.0
6:00 115.5 F 0.0
8:00 141.8 F 0.0
10:00 165.4 F 0.1
12:00 184.8 F 0.2
14:00 199.6 F 0.5
16:00 214.4 F 2.8
18:00 225.8 F 5.7
20:00 233.4 F 8.7
22:00 241.5 F 12.3
24:00 236.7 F 8.9
26:00 244.2 F 11.6
28:00 244.8 F 12.2
30:00 241.5 F 10.6
32:00 252.8 F 16.2
34:00 262.6 F 21.9
36:00 271.6 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.8
1:00 271.8 F 28.0
2:00 272.2 F 28.3
3:00 272.7 F 28.6
4:00 272.8 F 28.6

Min 271.6 F 27.7
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 201.0 F 0.0
10:00 222.0 F 0.0
15:00 242.9 F 0.0
20:00 238.3 F 0.0
25:00 229.8 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 36:49
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:15:14
CYCLE COMPLETE

12/16/2024 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____