
Midmark M11 - v1.0.0
Total Cycles: 1293

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.6 F 0.0
2:00 145.8 F 0.0
4:00 151.7 F 0.6
6:00 181.6 F 4.1
8:00 198.5 F 7.9
10:00 215.6 F 12.2
12:00 226.1 F 6.6
14:00 234.1 F 11.4
16:00 234.2 F 8.2
18:00 242.8 F 11.4
20:00 245.3 F 12.9
22:00 257.5 F 19.4
24:00 268.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.8 F 29.4
2:00 272.5 F 28.8
3:00 271.3 F 28.4
4:00 272.1 F 29.2

Min 271.2 F 28.3
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 235.0 F 0.0
10:00 224.2 F 0.0
15:00 225.4 F 0.0
20:00 227.6 F 0.0
25:00 231.2 F 0.0
30:00 201.2 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 25:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:03:47
CYCLE COMPLETE

10/22/2021 04:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____