
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001934
12:26 08/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:00
136.0 C 221kPa 3:00
136.3 C 225kPa 3:30
136.0 C 222kPa 4:00
136.3 C 223kPa 4:30
136.0 C 222kPa 5:00
136.3 C 225kPa 5:30
136.2 C 224kPa 6:00
136.4 C 226kPa 6:30
136.4 C 226kPa 6:30
VENTING 6:30
AIR DRYING 7:20
CYCLE COMPLETE 32:15

Digital Signature #
66B1953A36B06F38
