
Midmark M11 - v1.0.0
Total Cycles: 4208

Name: RITTER 1
Sterilizer ID: V2178198
01/16/2024 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.8 F 0.0
2:00 176.1 F 2.1
4:00 192.5 F 5.8
6:00 196.7 F 10.2
8:00 210.6 F 5.6
10:00 222.9 F 7.1
12:00 232.9 F 11.6
14:00 232.1 F 7.9
16:00 244.3 F 13.3
18:00 243.0 F 11.8
20:00 254.9 F 17.9
22:00 265.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 271.8 F 28.4
3:00 271.8 F 28.7
4:00 272.7 F 29.1

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 216.8 F 0.0
10:00 227.9 F 0.0
15:00 224.5 F 0.0
20:00 222.7 F 0.0
25:00 221.2 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:54
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:50
CYCLE COMPLETE

01/16/2024 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____