

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2850

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2023 09:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 186.2 F 2.1  
4:00 186.1 F 6.3  
6:00 204.5 F 11.7  
8:00 220.8 F 6.7  
10:00 231.7 F 12.1  
12:00 234.8 F 9.6  
14:00 237.2 F 9.4  
16:00 247.9 F 14.8  
18:00 259.9 F 21.4  
20:00 270.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.5 F 29.8  
2:00 271.7 F 29.0  
3:00 271.8 F 29.6  
4:00 272.4 F 29.7  
  
Min 271.2 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
5:00 208.3 F 0.0  
10:00 208.4 F 0.0  
15:00 200.5 F 0.0  
20:00 196.8 F 0.0  
25:00 214.2 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 20:50  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

09/08/2023 10:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_