
Midmark M11 - v1.0.0
Total Cycles: 1883

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 188.9 F 1.8
4:00 202.3 F 5.8
6:00 209.1 F 11.3
8:00 220.4 F 5.4
10:00 233.6 F 10.5
12:00 235.4 F 7.4
14:00 244.9 F 13.3
16:00 243.6 F 11.9
18:00 256.2 F 18.5
20:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 271.8 F 28.7
3:00 272.3 F 28.5
4:00 272.1 F 28.9

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 227.1 F 0.0
10:00 224.6 F 0.0
15:00 236.9 F 0.0
20:00 235.7 F 0.0
25:00 228.7 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

04/28/2022 09:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____