

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3051

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2023 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 71.3 F 0.0  
4:00 93.4 F 0.3  
6:00 107.0 F 2.4  
8:00 138.8 F 5.4  
10:00 160.8 F 9.3  
12:00 184.3 F 13.2  
14:00 201.0 F 5.6  
16:00 217.7 F 9.4  
18:00 228.8 F 11.9  
20:00 238.3 F 10.2  
22:00 234.5 F 8.5  
24:00 246.9 F 14.1  
26:00 258.7 F 20.5  
28:00 269.1 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.7  
1:00 271.4 F 29.6  
2:00 271.6 F 29.6  
3:00 272.1 F 29.7  
4:00 272.6 F 30.0

Min 271.2 F 29.6  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 235.8 F 0.0  
10:00 219.2 F 0.0  
15:00 185.7 F 0.0  
20:00 179.8 F 0.0  
25:00 178.4 F 0.0  
30:00 174.5 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:46  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:07:47  
CYCLE COMPLETE

10/25/2023 09:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_