
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001568
13:12 11/10/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:48
PRESSURIZING 6:46
STERILIZING 8:11
136.0 C 220kPa 8:11
136.1 C 220kPa 8:41
136.3 C 222kPa 9:12
136.5 C 220kPa 9:42
136.5 C 223kPa 10:12
136.4 C 222kPa 10:42
136.2 C 220kPa 11:12
136.3 C 222kPa 11:41
VENTING 11:42
AIR DRYING 12:49
CYCLE COMPLETE 22:18

Digital Signature #
7C5EF5B21F3AEFCA
