
Midmark M11 - v1.0.0
Total Cycles: 3953

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2023 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.2 F 0.0
2:00 137.4 F 1.9
4:00 148.4 F 5.0
6:00 172.6 F 8.7
8:00 211.0 F 13.5
10:00 213.6 F 6.2
12:00 225.6 F 10.8
14:00 229.2 F 6.8
16:00 241.6 F 12.4
18:00 239.8 F 10.6
20:00 252.8 F 16.9
22:00 264.0 F 23.5
24:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.1
2:00 271.3 F 28.4
3:00 272.4 F 29.3
4:00 272.0 F 28.8

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 207.2 F 0.0
10:00 230.9 F 0.0
15:00 229.2 F 0.0
20:00 229.3 F 0.0
25:00 229.5 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

10/27/2023 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____