

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001377  
15:47 05/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:03  
PRESSURIZING 4:04  
STERILIZING 5:14  
136.0 C 221kPa 5:14  
136.2 C 223kPa 5:44  
136.4 C 225kPa 6:14  
136.4 C 223kPa 6:44  
136.1 C 222kPa 7:14  
136.3 C 224kPa 7:44  
136.4 C 225kPa 8:14  
136.1 C 222kPa 8:44  
136.1 C 222kPa 8:44  
VENTING 8:44  
AIR DRYING 9:39  
CYCLE COMPLETE 11:40  
-----

Digital Signature #  
3AF65B390381863E  
-----