

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5474

Name: RITTER 1  
Sterilizer ID: V2178198  
04/03/2025 11:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.5 F 0.0  
2:00 152.8 F 2.6  
4:00 167.2 F 6.0  
6:00 195.7 F 10.0  
8:00 213.8 F 4.6  
10:00 220.6 F 7.8  
12:00 232.5 F 12.5  
14:00 232.7 F 8.8  
16:00 241.8 F 11.1  
18:00 243.7 F 12.6  
20:00 255.3 F 18.6  
22:00 265.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.6 F 29.6  
2:00 272.3 F 29.1  
3:00 271.5 F 28.9  
4:00 272.6 F 29.6

Min 271.4 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 192.6 F 0.0  
10:00 203.7 F 0.0  
15:00 219.1 F 0.0  
20:00 222.4 F 0.0  
25:00 225.1 F 0.0  
30:00 230.4 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:47  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:02:38  
CYCLE COMPLETE

04/03/2025 12:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_