

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001686  
10:55 24/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:48  
STERILIZING 3:48  
136.4 C 223kPa 3:48  
136.1 C 223kPa 4:18  
136.4 C 225kPa 4:48  
136.4 C 225kPa 5:18  
136.4 C 223kPa 5:48  
136.4 C 224kPa 6:18  
136.4 C 226kPa 6:48  
136.4 C 226kPa 7:18  
136.4 C 226kPa 7:18  
VENTING 7:18  
AIR DRYING 8:09  
CYCLE COMPLETE 11:46

-----  
Digital Signature #  
3C23EE84F888394C  
-----