
Midmark M11 - v1.0.5
Total Cycles: 3101

Name: RITTER2
Sterilizer ID: V2313228
11/08/2023 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.9 F 0.0
2:00 129.7 F 0.9
4:00 157.2 F 3.7
6:00 176.8 F 7.2
8:00 212.7 F 11.3
10:00 222.7 F 5.2
12:00 232.2 F 9.8
14:00 234.2 F 9.8
16:00 238.6 F 11.4
18:00 237.3 F 9.5
20:00 250.2 F 15.7
22:00 261.3 F 22.0
24:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 271.4 F 29.2
2:00 273.1 F 30.1
3:00 271.4 F 29.6
4:00 271.9 F 29.8

Min 271.0 F 29.1
Max 273.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 234.3 F 0.0
10:00 231.7 F 0.0
15:00 200.8 F 0.0
20:00 178.3 F 0.0
25:00 175.9 F 0.0
30:00 175.0 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

11/08/2023 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____