
Midmark M11 - v1.0.0
Total Cycles: 5457

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2025 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.0 F 0.0
2:00 140.9 F 2.1
4:00 147.9 F 5.5
6:00 173.3 F 9.4
8:00 213.7 F 9.8
10:00 218.0 F 5.1
12:00 222.4 F 9.3
14:00 239.4 F 11.7
16:00 237.9 F 10.1
18:00 237.8 F 8.9
20:00 247.3 F 14.0
22:00 258.6 F 20.1
24:00 269.0 F 26.7
26:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.6
2:00 272.7 F 29.5
3:00 272.2 F 29.0
4:00 271.5 F 28.8

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 190.8 F 0.0
10:00 191.9 F 0.0
15:00 194.9 F 0.0
20:00 199.7 F 0.0
25:00 211.5 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:08
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:04:56
CYCLE COMPLETE

03/27/2025 04:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____