
Midmark M11 - v1.0.5
Total Cycles: 1612

Name: RITTER2
Sterilizer ID: V2313228
10/18/2022 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 153.2 F 1.9
4:00 169.7 F 5.3
6:00 193.1 F 10.1
8:00 222.7 F 6.3
10:00 230.9 F 10.7
12:00 230.6 F 7.2
14:00 244.6 F 13.6
16:00 243.0 F 12.4
18:00 255.7 F 18.9
20:00 267.0 F 25.9
22:00 273.0 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 273.0 F 29.9
2:00 271.4 F 29.6
3:00 273.1 F 30.0
4:00 271.3 F 29.4

Min 271.2 F 29.3
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 231.9 F 0.0
10:00 227.0 F 0.0
15:00 226.1 F 0.0
20:00 171.2 F 0.0
25:00 162.0 F 0.0
30:00 157.4 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

10/18/2022 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____