
Midmark M11 - v1.0.0
Total Cycles: 894

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2021 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.8 F 0.0
2:00 140.4 F 1.4
4:00 164.3 F 4.2
6:00 187.6 F 8.1
8:00 209.7 F 12.8
10:00 218.5 F 7.5
12:00 231.1 F 12.6
14:00 234.1 F 9.5
16:00 236.5 F 9.0
18:00 246.7 F 14.1
20:00 258.9 F 20.7
22:00 270.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.3
2:00 272.1 F 29.5
3:00 272.0 F 29.1
4:00 272.3 F 29.6

Min 271.2 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 226.3 F 0.0
10:00 209.4 F 0.0
15:00 205.3 F 0.0
20:00 217.3 F 0.0
25:00 219.3 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:53
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

06/14/2021 02:59 PM
ACCEPTED: X
REJECTED:
Operator: U01