
Midmark M11 - v1.0.5
Total Cycles: 284

Name: RITTER2
Sterilizer ID: V2313228
08/18/2021 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.7 F 0.0
2:00 160.5 F 3.1
4:00 181.8 F 7.6
6:00 204.4 F 13.3
8:00 218.2 F 8.2
10:00 231.3 F 11.1
12:00 237.3 F 12.6
14:00 240.3 F 11.9
16:00 255.0 F 19.1
18:00 266.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.0
1:00 272.6 F 30.4
2:00 273.2 F 30.5
3:00 273.2 F 30.4
4:00 272.5 F 30.4

Min 271.9 F 29.9
Max 273.5 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 241.0 F 0.0
10:00 228.7 F 0.0
15:00 223.6 F 0.0
20:00 208.2 F 0.0
25:00 203.4 F 0.0
Dry Cycle Aborted

FILLING: 1:28
HEATING: 19:36
STERILIZING: 4:00
VENTING: 2:56
DRYING: 26:26
TOTAL CYCLE: 00:54:26

08/18/2021 01:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____