

-----  
-----  
Midmark M11 - v1.0.0  
Total Cycles: 4757

Name: RITTER 1  
Sterilizer ID: V2178198  
07/22/2024 04:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.8 F 0.0  
2:00 168.2 F 3.8  
4:00 185.3 F 8.3  
6:00 206.8 F 12.8  
8:00 220.1 F 8.0  
10:00 233.5 F 13.5  
12:00 237.0 F 10.7  
14:00 235.9 F 9.0  
16:00 250.8 F 16.0  
18:00 263.1 F 23.1  
20:00 271.5 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.1  
2:00 271.8 F 29.0  
3:00 272.4 F 29.0  
4:00 271.6 F 28.9

Min 271.3 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 202.1 F 0.0  
10:00 207.6 F 0.0  
15:00 204.1 F 0.0  
20:00 203.4 F 0.0  
25:00 203.5 F 0.0  
30:00 205.0 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

07/22/2024 05:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_