
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000568
13:48 06/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:28
STERILIZING 4:45
136.0 C 222kPa 4:45
136.2 C 230kPa 5:15
136.1 C 230kPa 5:45
136.0 C 231kPa 6:15
136.4 C 233kPa 6:45
136.4 C 232kPa 7:15
136.4 C 233kPa 7:45
136.5 C 233kPa 8:15
136.5 C 233kPa 8:15
VENTING 8:15
AIR DRYING 9:11
CYCLE COMPLETE 12:57

Digital Signature #
6B31DA2399D51FCA
