

-----  
Midmark M11 - v1.0.5  
Total Cycles: 879

Name: RITTER2  
Sterilizer ID: V2313228  
03/14/2022 01:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
2:00 190.6 F 4.0  
4:00 203.2 F 9.4  
6:00 221.4 F 7.8  
8:00 231.2 F 11.7  
10:00 234.8 F 9.8  
12:00 236.8 F 8.8  
14:00 249.6 F 15.8  
16:00 261.7 F 22.9  
18:00 272.9 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.4  
1:00 272.3 F 29.9  
2:00 271.7 F 29.5  
3:00 272.3 F 29.9  
4:00 271.6 F 29.2  
  
Min 271.6 F 29.2  
Max 272.4 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 199.9 F 0.0  
10:00 210.2 F 0.0  
15:00 212.6 F 0.0  
20:00 211.4 F 0.0  
25:00 210.3 F 0.0  
30:00 209.7 F 0.0  
Drying Complete

FILLING: 2:26  
HEATING: 18:20  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

03/14/2022 02:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_