
Midmark M11 - v1.0.0
Total Cycles: 3371

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 181.5 F 0.0
4:00 192.4 F 3.6
6:00 211.9 F 8.8
8:00 224.5 F 9.0
10:00 233.2 F 10.0
12:00 235.1 F 8.5
14:00 245.3 F 13.4
16:00 244.0 F 12.3
18:00 256.6 F 18.9
20:00 267.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.9
2:00 271.6 F 28.3
3:00 272.7 F 29.0
4:00 271.8 F 28.4

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 222.3 F 0.0
10:00 230.8 F 0.0
15:00 228.0 F 0.0
20:00 223.7 F 0.0
25:00 221.7 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

06/12/2023 10:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____