
Midmark M11 - v1.0.0
Total Cycles: 3653

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2023 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 189.3 F 3.0
4:00 187.9 F 6.9
6:00 206.9 F 11.9
8:00 221.4 F 6.1
10:00 233.5 F 11.2
12:00 232.8 F 8.2
14:00 242.4 F 11.5
16:00 244.9 F 12.8
18:00 257.2 F 19.3
20:00 268.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.6 F 28.5
2:00 272.0 F 28.7
3:00 272.4 F 29.1
4:00 271.9 F 29.0

Min 271.3 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 210.3 F 0.0
10:00 211.6 F 0.0
15:00 215.7 F 0.0
20:00 217.3 F 0.0
25:00 225.0 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

08/15/2023 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____