
Midmark M11 - v1.0.0
Total Cycles: 2194

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2022 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 179.3 F 3.7
4:00 187.1 F 8.0
6:00 204.4 F 12.6
8:00 216.9 F 6.4
10:00 232.2 F 11.3
12:00 231.5 F 7.8
14:00 243.9 F 12.7
16:00 243.4 F 12.1
18:00 255.9 F 18.6
20:00 266.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 272.0 F 28.6
3:00 271.6 F 28.7
4:00 272.6 F 29.2

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 238.8 F 0.0
10:00 224.4 F 0.0
15:00 217.5 F 0.0
20:00 220.4 F 0.0
25:00 228.3 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

08/03/2022 02:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____