
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005819
8:49 23/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:34
STERILIZING 3:43
136.5 C 224kPa 3:43
136.3 C 225kPa 4:13
136.4 C 222kPa 4:43
136.4 C 223kPa 5:13
136.4 C 224kPa 5:43
136.3 C 222kPa 6:13
136.4 C 224kPa 6:43
136.4 C 223kPa 7:13
136.4 C 223kPa 7:13
VENTING 7:13
AIR DRYING 8:03
CYCLE COMPLETE 8:08

Digital Signature #
6DF00F95BF882CFC
