
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003890
12:41 27/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:42
136.4 C 223kPa 3:42
136.2 C 219kPa 4:12
136.3 C 224kPa 4:42
136.4 C 223kPa 5:12
136.4 C 224kPa 5:42
136.2 C 222kPa 6:12
136.3 C 224kPa 6:42
136.4 C 223kPa 7:12
136.4 C 223kPa 7:12
VENTING 7:13
AIR DRYING 8:02
CYCLE COMPLETE 9:46

Digital Signature #
44AA7DD7A3BFDCF6
