
Midmark M11 - v1.0.0
Total Cycles: 5243

Name: RITTER 1
Sterilizer ID: V2178198
01/14/2025 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.1 F 0.0
2:00 100.9 F 0.0
4:00 121.4 F 0.0
6:00 156.3 F 0.3
8:00 182.9 F 2.5
10:00 196.1 F 5.6
12:00 206.3 F 9.4
14:00 216.2 F 13.3
16:00 226.2 F 7.4
18:00 238.8 F 12.0
20:00 235.8 F 9.0
22:00 240.3 F 10.3
24:00 246.6 F 13.5
26:00 258.4 F 19.8
28:00 269.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.8
2:00 271.4 F 28.2
3:00 272.5 F 28.9
4:00 271.6 F 28.2

Min 271.4 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 202.9 F 0.0
10:00 186.4 F 0.0
15:00 188.9 F 0.0
20:00 209.8 F 0.0
25:00 226.2 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 29:07
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:07:31
CYCLE COMPLETE

01/14/2025 05:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____