
Midmark M11 - v1.0.5
Total Cycles: 611

Name: RITTER2
Sterilizer ID: V2313228
12/08/2021 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.8 F 0.0
2:00 176.0 F 3.0
4:00 180.4 F 7.5
6:00 203.5 F 13.2
8:00 225.4 F 7.6
10:00 237.3 F 13.1
12:00 239.3 F 11.2
14:00 238.7 F 10.2
16:00 253.0 F 17.3
18:00 265.2 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.6 F 29.7
2:00 271.6 F 28.8
3:00 272.7 F 29.6
4:00 271.6 F 28.8

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 222.2 F 0.0
10:00 245.6 F 0.0
15:00 218.8 F 0.0
20:00 211.1 F 0.0
25:00 213.3 F 0.0
30:00 213.7 F 0.0

Drying Complete

FILLING: 1:31
HEATING: 19:37
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

12/08/2021 11:41 AM
ACCEPTED: X
REJECTED:
Operator: U02