
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000224
7:55 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:43
PRESSURIZING 4:43
STERILIZING 5:54
136.0 C 222kPa 5:54
136.0 C 230kPa 6:24
136.2 C 231kPa 6:54
136.3 C 232kPa 7:24
136.2 C 229kPa 7:54
136.4 C 233kPa 8:24
136.2 C 230kPa 8:54
136.1 C 230kPa 9:24
136.1 C 230kPa 9:24
VENTING 9:24
AIR DRYING 10:26
CYCLE COMPLETE 10:30

Digital Signature #
2040937F0AA6C333
