

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1101

Name: RITTER2  
Sterilizer ID: V2313228  
05/12/2022 08:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.1 F 0.0  
2:00 166.7 F 2.2  
4:00 167.6 F 6.0  
6:00 191.8 F 11.3  
8:00 223.9 F 6.0  
10:00 234.3 F 11.4  
12:00 234.1 F 8.6  
14:00 240.1 F 10.7  
16:00 246.3 F 13.6  
18:00 258.8 F 20.5  
20:00 270.0 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 271.8 F 29.6  
2:00 272.4 F 29.6  
3:00 272.4 F 29.5  
4:00 272.7 F 29.4

Min 271.4 F 29.2  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 235.9 F 0.0  
10:00 219.4 F 0.0  
15:00 221.4 F 0.0  
20:00 219.0 F 0.0  
25:00 232.2 F 0.0  
30:00 190.1 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:37  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

05/12/2022 09:24 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02