
Midmark M11 - v1.0.0
Total Cycles: 1358

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2021 07:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE
WATER LOW
CHAMBER TEMP 72F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

11/11/2021 07:54 AM