

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3171

Name: RITTER 1  
Sterilizer ID: V2178198  
04/27/2023 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.8 F 0.0  
2:00 192.0 F 3.2  
4:00 218.9 F 8.5  
6:00 230.1 F 9.5  
8:00 235.7 F 10.0  
10:00 237.7 F 8.0  
12:00 243.0 F 11.5  
14:00 247.0 F 13.4  
16:00 259.9 F 20.6  
18:00 271.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 271.9 F 28.4  
2:00 272.5 F 28.9  
3:00 271.4 F 28.3  
4:00 272.2 F 28.6  
  
Min 271.4 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 232.6 F 0.0  
10:00 227.8 F 0.0  
15:00 224.3 F 0.0  
20:00 224.2 F 0.0  
25:00 225.5 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 18:32  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

04/27/2023 04:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_