
Midmark M11 - v1.0.0
Total Cycles: 836

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2021 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 158.0 F 1.6
4:00 169.7 F 3.9
6:00 194.2 F 7.5
8:00 209.7 F 11.4
10:00 217.3 F 5.0
12:00 229.6 F 9.7
14:00 234.8 F 9.9
16:00 240.5 F 11.6
18:00 238.7 F 9.9
20:00 251.5 F 16.1
22:00 262.4 F 22.4
24:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 272.2 F 28.7
3:00 271.4 F 28.4
4:00 272.3 F 29.2

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 208.5 F 0.0
10:00 201.1 F 0.0
15:00 200.2 F 0.0
20:00 198.9 F 0.0
25:00 204.6 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:30
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:03
CYCLE COMPLETE

05/26/2021 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____