
Midmark M11 - v1.0.0
Total Cycles: 1157

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2021 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.2 F 0.0
2:00 152.5 F 1.0
4:00 171.0 F 4.6
6:00 196.0 F 9.0
8:00 212.8 F 7.6
10:00 226.6 F 8.3
12:00 241.0 F 12.3
14:00 239.8 F 11.0
16:00 237.3 F 9.3
18:00 251.5 F 16.1
20:00 263.4 F 22.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.6 F 28.9
2:00 272.6 F 29.2
3:00 271.5 F 28.4
4:00 272.5 F 28.8

Min 271.5 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 220.1 F 0.0
10:00 220.2 F 0.0
15:00 206.0 F 0.0
20:00 200.4 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:04
DRYING: 24:31
TOTAL CYCLE: 00:54:28

09/09/2021 02:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____