
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001394
13:48 21/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 4:25
STERILIZING 5:45
136.0 C 220kPa 5:45
136.4 C 224kPa 6:15
136.4 C 225kPa 6:45
136.1 C 223kPa 7:15
136.4 C 225kPa 7:45
136.4 C 224kPa 8:15
136.1 C 222kPa 8:45
136.3 C 224kPa 9:15
136.3 C 224kPa 9:15
VENTING 9:15
AIR DRYING 10:15
CYCLE COMPLETE 25:15

Digital Signature #
AD9633AF2AF18C49
