
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001193
8:09 28/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:09
PRESSURIZING 4:11
STERILIZING 5:25
136.0 C 220kPa 5:25
136.4 C 224kPa 5:55
136.4 C 224kPa 6:25
136.3 C 224kPa 6:55
136.2 C 222kPa 7:25
136.4 C 224kPa 7:55
136.4 C 221kPa 8:25
136.1 C 222kPa 8:55
136.1 C 222kPa 8:55
VENTING 8:55
AIR DRYING 9:49
CYCLE COMPLETE 24:49

Digital Signature #
C9ADB72A58CFB1A8
