
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001559
16:13 01/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:04
136.0 C 220kPa 3:04
136.3 C 223kPa 3:34
136.2 C 223kPa 4:04
136.3 C 225kPa 4:34
136.1 C 224kPa 5:04
136.4 C 224kPa 5:34
136.4 C 225kPa 6:04
136.4 C 224kPa 6:34
136.4 C 224kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 37:26

Digital Signature #
7A34C19CFB54EE2C
